

# Temporal Artery Thermometry

*Three good reasons for hospital wide use*

## 1. Lowest Cost System

Probe covers not required - available if needed

- Reduce thermometry cost by 90%
- Reduce associated waste by 90%
- Lifetime Warranty

## 2. Highest Accuracy



Accuracy supported by more than 40 peer reviewed published studies

## 3. Most Liked by Patients

- Most comfortable thermometer
- Least invasive thermometer
- Home models available at retailers



Virtual Classroom Training

# Accuracy and Benefits of Using Exergen Temporal Artery Thermometry

## Frequently Asked Questions at Products and Value Analysis Committee Meetings

1. **Settled Science:** It is accurate to report that Exergen Temporal Artery Thermometers have never been proven inaccurate when compared to the true Gold Standards of pulmonary artery catheters, esophageal probes, rectal temperature on a stable patient, and outcome. When temperature is rapidly changing, however, rectal temperature is well known to lag behind arterial temperature on adults, the scientific evidence for which dates back to the late 1960's with the invention of the PA catheter, and, most recently, rectal lag has been identified in infants (19).<sup>1</sup>
2. **What are the benefits for our hospital in changing to Exergen TA Thermometry?**

The significant clinical and fiscal benefits of the Exergen temporal artery thermometers have been confirmed by about 60% of the hospitals throughout the United States:

  - a. Clinically, the TA thermometers are as accurate as the most invasive methods.
  - b. Fiscally, the method is 90% less expensive than any other method of thermometry, and provides a payback in approximately 8 months, with near zero expenditure from that point forward.
3. **How do the results compare to rectal temps, oral temps, and ear temps?**
  - a. Several peer reviewed published studies conducted in major university hospitals on newborn infants through geriatric patients have demonstrated the temperatures measured by the Exergen Temporal Artery Thermometer to be: significantly more accurate than ear thermometers (6,10,14,15,16,18,21,23,24,25,26,28,30,39); as accurate as rectal temperature (1,2,4,7,8,10,11,12,15,18,19,27,31,40,42,44), core temperature as measured with an esophageal probe, pulmonary artery catheter, or brain thermistor (1,2,3,6,9,11,12,22,23,25,26,30), and outcome (5,19,25,28,32,38,41,43).
4. **At what age is the method validated?**
  - a. Newborns 0-3mo (7,8,11,13,18,19,20,27,38,40,41,42), infants and children (1,2,4,7,8,9,18,19,22,31,38,40,41,42,44), adults through geriatric patients (3,5,6,9,12,14,15,16,17,23,24,25,26,28,30,37).
5. **Below the age group the thermometer is validated, what equipment will we use and how will that equipment be supported?**
  - a. Age is not a factor as the TA thermometer has been demonstrated to be accurate for use on all ages.
  - b. Nonetheless, when the patient is under a radiant warmer or in an isolette, the TAT should be allowed to equilibrate ~20 minutes before use. If not possible, Exergen recommends using the Exergen LXN Instant infrared axillary thermometer, an instrument successfully used for many years, mainly in Neonatal Intensive Care Units. Studies available from Exergen.
6. **What is the evidence that we will not spread disease with this device?**
  - a. The TA thermometer features a silver ion antimicrobial head, unique in thermometry, but widely used clinically in such items as catheters, stents, wound dressings, spinal implants, IV flow valves, in which the silver ions protects by naturally and continuously resisting the growth of microbes.
  - b. Laboratory tests demonstrate the controlled release of silver ions from the sensor head provides continuous antimicrobial protection for more than five years. The antimicrobial compound works proactively against a broad spectrum of bacteria, fungi and other microbes. In laboratory studies, this has been proven to reduce bacteria on the treated product by as much as 99.999%, or 5-log reduction. Studies available from Exergen.
  - c. The TA thermometer is used by ~60% of the hospitals in the US. Of these, ~92% have approved wiping the sensorhead between patients with an alcohol swab, or with whatever has been approved in their facility for wiping the stethoscope diaphragm between patients.
  - d. Resposable caps and sheaths that enclose the entire instrument are available for use where preferred.

---

<sup>1</sup> Numbered references in parenthesis refer to attached list of "Peer-Reviewed Papers and Abstracts on Exergen Temporal Artery Thermometers." Additional clinical information can be found at Exergen's Clinical Website at [www.TAThermometry.org](http://www.TAThermometry.org).

## Peer-Reviewed Published Papers, Abstracts, Letters on Exergen Temporal Artery Thermometry

1.	Al-Mukhaizeem F, Allen U, Komar L, et al (University of Toronto/Hospital for Sick Children). Validation of the temporal artery thermometry by its comparison with the esophageal method in children. Pediatric Academic Societies Annual Meeting, May 3-6, 2003, Seattle, WA
2.	Al-Mukhaizeem F, Allen U, Komar L, et al (University of Toronto/Hospital for Sick Children). Comparison of temporal artery, rectal and esophageal core temperatures in children: Results of a pilot study. <i>Journal of Pediatric and Child Health</i> , Vol 9, No 7, pp 461-465, 2004
3.	Artz BA, March KS, Grim RD ( WellSpan Health–York Hospital). Clinical Nurse Specialists empowering staff to improve patient outcomes in temperature measurement: from PI/EBP to nursing research. 2011 National Association of Clinical Nurse Specialists National Conference Abstracts, March 10-12, 2011, Baltimore MD
4.	Bahorski J, Repasky T, Ranner D, Fields A, Jackson M, Moultry L, Pierce K, Sandell M (Tallahassee Memorial Healthcare). Temperature measurement in pediatrics: a comparison of the rectal method versus the temporal artery method. In Press, Corrected Proof, Available online 24 February 2011, <i>Journal of Pediatric Nursing</i> (2011).
5.	Barringer LB, Evans CW, Ingram LL, Tisdale PP, Watson SP, Janken JK (Presbyterian Hospital Matthews). Agreement between temporal artery, oral, and axillary temperature measurements in the perioperative period. <i>J Perianesth Nurs</i> . 2011 Jun;26(3):143-50.
6.	Bridges E, Thomas K (University of Washington). Noninvasive measurement of body temperature in critically ill patients. <i>Crit. Care Nurse</i> . 2009; 29(3): p. 94-97
7.	Burdjalov VF, Combs A, Nachman S, Baumgart S (SUNY at Stony Brook). Non-Invasive infrared temperature assessment of the temporal artery for core temperature determination in premature neonates, Presented <i>American Pediatric Society and the Society for Pediatric Research</i> , May 1, 2001.
8.	Callanan D (Christus Santa Rosa Children's Hospital). Detecting fever in young infants: reliability of perceived, pacifier, and temporal artery temperatures in infants younger than 3 months of age. <i>Pediatr Emerg Care</i> . 2003 Aug;19(4):240-3.
9.	Calonder EM, Sendelbach S, Hodges JS, Gustafson C, Machemer C, Johnson D, Reiland L (Abbott Northwestern Hospital). Temperature measurement in patients undergoing colorectal surgery and gynecology surgery: a comparison of esophageal core, temporal artery, and oral methods. <i>Journal of PeriAnesthesia Nursing</i> , Volume 25, Issue 2, April 2010, Pages 71-78
10.	Canales AE (Texas Tech University Health Sciences Center). OTC device: temporal scanner TAT-2000C. <i>J Am Pharm Assoc</i> (Wash DC). 2007 Jan-Feb;47(1):112.
11.	Carr EA, Wilmoth ML, Eliades AB, Baker PJ, Shelestak D, Heisroth KL, Stoner KH (Akron Children's Hospital). Comparison of Temporal Artery to Rectal Temperature Measurements in Children Up to 24 Months, <i>Journal of Pediatric Nursing</i> , In Press, [Epub ahead of print], Jan 25, 2010.
12.	Carroll D, Finn C, Gill S, et al (Massachusetts General Hospital). A comparison of measurements from a temporal artery thermometer and a pulmonary artery catheter thermometer. <i>Am J Crit Care</i> . 2004;13:258.
13.	Chiu SH, Anderson GC, Burkhammer MD (University of Akron/Case Western Reserve University). Newborn temperature during skin-to-skin breastfeeding in couples having breastfeeding difficulties. <i>Birth</i> . 2005 Jun;32(2):115-21.
14.	Dybwik K, Nielsen EW. Infrared temporal temperature measurement. <i>Journal of the Norwegian Medical Association</i> 2003; 123: 3025-6.
15.	Espenhein A (County Hospital in Herlev, Denmark). Temporal temperature measurement. <i>Sygeplejersken</i> 2006;(17):50-2.
16.	Fetzer SJ, Lawrence A (Southern New Hampshire Medical Center). Tympanic membrane versus temporal artery temperatures of adult perianesthesia patients. <i>J Perianesth Nurs</i> . 2008 Aug;23(4):230-6.
17.	Foy S, McGillicuddy D, Pompei F, Sanchez L (Beth Israel Medical Center, Boston MA). Body Temperature Surveillance and Reporting in the Emergency Department: A Practical Sentinel for Pandemics and Bioterrorism. Presented at Society for Academic Emergency Medicine Annual Meeting, Phoenix AZ , June 3-6, 2010.
18.	Greenes DS, Fleisher GR. (Boston Childrens Hospital and Harvard Medical School). Accuracy of a noninvasive temporal artery thermometer for use in infants. <i>Arch Pediatr Adolesc Med</i> , Vol 155, pp 376-381, Mar 2001
19.	Greenes DS, Fleisher GR. (Boston Childrens Hospital and Harvard Medical School). When body temperature changes, does rectal temperature lag? <i>Journal of Pediatrics</i> , 02.037, pp 824-826, September 2004.
20.	Gunawan M, Soetjningsih I ( Udayana University, Sanglah Hospital, Denpasar, Indonesia). Comparison of the accuracy of body temperature measurements with temporal artery thermometer and axillary mercury thermometer in term newborns. <i>Paediatr Indones</i> , Vol. 50, No. 2, March 2010.
21.	Harper CM ( Royal Sussex County Hospital Brighton, UK). The need for an accurate noninvasive thermometer. <i>Anesth Analg</i> . 2009 Jul;109(1):288; author reply 288-9.

22.	Hebbar K, Fortenberry JD, Rogers K, Merritt R, Easley K. (Children's Healthcare of Atlanta at Egleston). Comparison of temporal artery thermometer to standard temperature measurements in pediatric intensive care unit patients. <i>Pediatr Crit Care Med.</i> 2005 Sep;6(5):557-61.
23.	Kirk D, Rainey T, Vail A, Childs C (University of Manchester, Salford Royal Foundation Trust). Infra-red thermometry: the reliability of tympanic and temporal artery readings for predicting brain temperature after severe traumatic brain injury. <i>Crit Care.</i> 2009 May 27;13(3):R81. [Epub ahead of print]
24.	Langham GE, Maheshwari A, Contrera K, You J, Mascha E, Sessler DI (Case Western Reserve University). Noninvasive temperature monitoring in postanesthesia care units. <i>Anesthesiology</i> , V 111, No 1, Jul 2009
25.	Lawson L, Bridges E, Ballou I, Eraker R, Greco S, Shively J, Sochulak V. (University of Washington). Temperature measurement in critically ill adults. <i>Am. J. Crit. Care.</i> , May 2006; 15: 324 - 346.
26.	Lawson L, Bridges E, Ballou I, Eraker R, Greco S, Shively J, Sochulak V. (University of Washington). Accuracy and precision of noninvasive temperature measurement in adult intensive care patients. <i>Am. J. Crit. Care.</i> , Sep 2007; 16:5, 485-496.
27.	Lee G, Flannery-Bergey D, Randall-Rollins K, Curry D, Rowe S, Teague M, Tuininga C, Schroeder S (Exempla Lutheran Medical Center). Accuracy of temporal artery thermometry in neonatal intensive care infants. <i>Advances in Neonatal Care</i> , Vol. 11, No. 1, pp. 62-70, Feb 2011.
28.	Makic MB, VonRueden KT, Rauhen CA, Chadwick J. Evidence-based practice habits: putting more sacred cows out to pasture. <i>Crit Care Nurse.</i> 2011 Apr;31(2):38-61; quiz 62.
29.	Martinez EA, Krenzschek D, Hobson D, Hunt D (Johns Hopkins Medical Institutions). The structure and processes of care delivery impact postoperative normothermia. <i>Anesthesiology</i> 2007; 107: A496.
30.	Myns D, DeWaele J, Defloor T, Blot S, Colardyn F (Ghent University Hospital, Ghent, Belgium). Temporal scanner thermometry: a new method of core temperature measurement in intensive care patients. <i>SMJ</i> 2005 45(1): 15-18.
31.	Paul IM, Sturgis SA, Yang C, Engle L, Watts H, Berlin CM Jr (Penn State College of Medicine). Efficacy of standard doses of ibuprofen alone, alternating, and combined with acetaminophen for the treatment of febrile children. <i>Clin Ther.</i> 2010 Dec;32(14):2433-40.
32.	Pittman R and Waters R (CaroMont Health Care, Gastonia, NC). Do our patients have hypothermia? Temporal versus oral thermometers. <i>Journal of PeriAnesthesia Nursing</i> Volume 24, Issue 3, June 2009, Page e18.
33.	Pompei F, Pompei M. Non-invasive temporal artery thermometry: Physics, Physiology, and Clinical Accuracy, presented at <i>Medical Thermometry for SARS Detection, SPIE Defense and Security Symposium</i> , available in <i>Conference Proceedings</i> , April, 2004.
34.	Pompei F. Insufficiency in thermometer data. <i>Anesth Analg.</i> 2003 Mar;96(3):908-9.
35.	Pompei F. RE: A brief report on the normal range of forehead temperature as determined by noncontact, handheld, infrared thermometer. <i>Am J Infect Control.</i> 2006 May;34(4):248-9.
36.	Pompei F. Misguided guidelines on noninvasive thermometry. <i>Crit Care Med.</i> 2009 Jan;37(1):383; author reply 383-4.
37.	Routhier D, Hostler D, Wolfson A, Wheeler M, Reynolds J (University of Pittsburgh). Comparison of temporal artery and oral temperatures in the emergency department. <i>ACAD EMERG MED</i> , May 2006, Vol. 13, No. 5, Suppl. 1, <a href="http://www.aemj.org">www.aemj.org</a> , p. S99
38.	Roy S, Powell K, Gerson LW (Akron Children's Hospital). Temporal artery temperature measurements in healthy infants, children, and adolescents. <i>Clinical Pediatrics</i> , pp 433-437, June 2003.
39.	Sandlin D (Southern Hills Medical Center, Nashville TN). New Product Review: Temporal Artery Thermometry, <i>Journal of PeriAnesthesia Nursing</i> , Vol. 18, No 6 (December) 2003, pp 419-421.
40.	Schuh S, Komar L, Stephens D, Chu L, Read S, Allen U (University of Toronto/Hospital for Sick Children). Comparison of the temporal artery and rectal thermometry in children in the emergency department. <i>Pediatric Academic Societies Annual Meeting</i> , May 3-6, 2003, Seattle, WA.
41.	Schuh S, Komar L, Stephens D, Chu L, Read S, Allen U (University of Toronto/Hospital for Sick Children). Comparison of the temporal artery and rectal thermometry in children in the emergency department. <i>Pediatric Emergency Care</i> , Vol 20, No. 11, Nov 2004
42.	Siberry GK, Diener-West M, Schappell E, Karron RA (Department of Pediatrics, School of Medicine, The Johns Hopkins University). Comparison of temple temperatures with rectal temperatures in children under two years of age. <i>Clinical Pediatrics</i> , pp 405-414, July/August 2002.
43.	Szmuk P, Curry BP, Sheeran PW, Farrow-Gillespie AC, Ezri T (UT Southwestern and Children's Medical Center, Dallas, Texas). Perioperative temperature audit in a large pediatric hospital. <i>Anesthesiology</i> 2007; 107: A1612.
44.	Titus MO, Hulsey T, Heckman J, Losek JD (Medical University of South Carolina and Children's Hospital). Temporal artery thermometry utilization in pediatric emergency care. <i>Clinical Pediatrics</i> , Mar 2009; vol. 48: pp. 190 - 193.