

Basics of Using Exergen Temporal Scanner Thermometers




Brush hair away if covering the forehead or ear



Place probe FLUSH on forehead, depress button & keep depressed until you are done



3 Slide STRAIGHT across forehead, NOT down the side of the face



IMPORTANT Dirty Lens = Low Temps

Clean lens in center of probe every 2 weeks with a Q-TIP dampened with an alcohol prep pad.



IMPORTANT Core Temp is about 99.4° (37.4°C)

Temporal *Scanner* measures CORE temperature, about 1°F (0.6°C) higher than oral temperature.

Remove, read and record temperature





Watertown, MA 02472 USA • www.exergen.com • www.TAthermometry.org

PN: 818647A ©2009 Exergen Corporation