



## EU DECLARATION OF CONFORMITY



**Manufacturer Name:** Gems Sensors and Controls  
**Manufacturer Address:** Lennox Road, Basingstoke, Hampshire, RG22 4AW  
**Product Type:** **2200 Series**  
Strain Gauge Pressure Transducers and Pressure Transmitters  
Full Scale Pressure Range -1 Bar to 400 Bar  
Operating Temperature Range -40°C to +125°C

**Pressure Equipment Directive:** 2014/68/EU  
Equipment marked CE2797 is classed as a safety accessory and can be used as a safety-related device on Category IV pressure equipment

**Conformity Assessment Procedure:** CE2797: Modules B + D CE: Sound Engineering Practice

**Module B EC Type Examination Certificate reference(s):** CE72108

**Compliance with other Directives:**  
**EMC Directive:** 2014/30/EU  
**Harmonised Standards Used:** BS EN 61326-1:2013 BS EN 61326-2-3:2013

**RoHS3 Directive** 2015/863/EU



### Notified Bodies:

**Notified Body for PED Module B Assessment:**  
BSI 2797  
BSI Group The Netherlands B.V.,  
Registered in the Netherlands  
under number 33264284,  
at John M. Keynesplein 9, 1066 EP  
Amsterdam, The Netherlands  
A member of BSI Group of  
Companies

**Notified Body for PED Module D Assessment:**  
BSI 2797  
BSI Group The Netherlands B.V.,  
Registered in the Netherlands  
under number 33264284,  
at John M. Keynesplein 9, 1066 EP  
Amsterdam, The Netherlands  
A member of BSI Group of  
Companies

**This apparatus must not be put into service until the equipment into which it is to be incorporated has been declared in conformity with the provisions of the relevant New Approach Directive**

On behalf of Gems Sensors and Controls, I declare that on the date the product(s) listed on this declaration is placed on the market, the product(s) conforms with all technical and regulatory requirements of the above listed directives.

  
\_\_\_\_\_  
(signature)  
  
\_\_\_\_\_  
(date)

Steve Allen  
Global Operations Director