



CERTIFICATE No. 51909  
 CONTRACT No. INV 100202

## SPECIALIST WEIGHING COUNTING AND PROCESS SOLUTIONS

Customer Name  
 Address  
**MAY ECO LIMITED**  
**UNIT 3B HOLLOWAY DRIVE**  
**HARDLEY INDUSTRIAL ESTATE**  
**MANCHESTER**  
**M28 2LA**

B.S.I. Registered Firm No. FM11336  
 9938/9, 1318, 9963, 9981, 3528, 10068/9, 3697  
 10009, 10173, 3650/1, 9964, 3654/5, 3695, 3720

TEST WEIGHT CERTIFICATE OF COMPLIANCE No.

Contact **DOM O'BRIEN** Date of Calibration **Oct 2023** Engineers Signature

| PLANT NUMBER         | LOCATION | MAKE    | TYPE     | SERIAL No. | Capacity | Smallest Increment (1 Div) | ACCURACY ON INSPECTION |          |          | ACCURACY AFTER CALIBRATION |                    |          | MACHINE OUT OF TOLERANCE REPAIR RECOMMENDATIONS |
|----------------------|----------|---------|----------|------------|----------|----------------------------|------------------------|----------|----------|----------------------------|--------------------|----------|---|
|                      |          |         |          |            |          |                            | Measurement Points     | 25% Load | 50% Load | Full Load                  | Measurement Points | 25% Load |   |
| 1                    | WORKS    | MONSTER | PLATFORM | N/A        | 1500Kg   | 500g                       | ✓                      | ✓        |          | ✓                          | ✓                  |          |   |
| 1 REPEATABILITY TEST |          |         |          |            |          |                            | ✓                      | ✓        |          | ✓                          | ✓                  |          |   |
| 2 REPEATABILITY TEST |          |         |          |            |          |                            |                        |          |          |                            |                    |          |   |
| 3 REPEATABILITY TEST |          |         |          |            |          |                            |                        |          |          |                            |                    |          |   |
| 4 REPEATABILITY TEST |          |         |          |            |          |                            |                        |          |          |                            |                    |          |   |

ECCENTRIC LOAD ACCURACY

|   |   |   |  |  |  |  |  |  |  |
|---|---|---|--|--|--|--|--|--|--|
| 1 | ✓ | ✓ |  |  |  |  |  |  |  |
| 2 | ✓ | ✓ |  |  |  |  |  |  |  |
| 3 |   |   |  |  |  |  |  |  |  |
| 4 |   |   |  |  |  |  |  |  |  |

✓ INDICATES ACCEPTANCE WITH TOLERANCES OF 1 DIV. UP TO 500 DIVS. 2 DIVS. 500-2000 DIVS. & 3 DIVS. ABOVE 2000 DIVS.

This is to certify that the above equipment has been tested using weights that are traceable to National Standards



Our quality system is accredited to BS EN ISO 9001  
 Certificate Number FM11336